



## Reconvolution

Reconvolution is used in time domain data analysis for removal of broadening effects due to instrumental resolution. In contrast to [deconvolution](#), data processing using reconvolution techniques does not aim at retrieving the raw data as seen by an ideal instrument, but does usually compare artificial decays (simulated on the base of a model) convolved with the measured [IRF](#).

Reconvolution is used in PicoQuant software packages (mainly [FluoFit](#) and [SymPhoTime](#)) for compensating IRF effects.

Copyright of this document belongs to PicoQuant GmbH. No parts of it may be reproduced, translated or transferred to third parties without written permission of PicoQuant GmbH. All information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice.



PicoQuant GmbH  
Rudower Chaussee 29 (IGZ)  
12489 Berlin  
Germany

P +49-(0)30-1208820-89  
F +49-(0)30-1208820-90  
info@picoquant.com  
www.picoquant.com